


EPA Method 8260C - Low-Level Volatiles by Full-Scan GC/MS


CHEMICAL OF CONCERN	CAS	MATRIX	GROUND WATER ^{ING} TIER 1 PCL (PPB)	MDL (PPB)	PQL (PPB)
Ethylene Dibromide (EDB) 1,2-Dibromoethane	106-93-4	Aqueous	0.05	0.01	0.03
1,2-Dibromo-3-Chloropropane (DCBP)	96-12-8	Aqueous	0.20	0.05	0.10
1,2,3-Trichloropropane	96-18-4	Aqueous	0.068	0.03	0.05

CHEMICAL OF CONCERN	CAS	MATRIX	^{GW} SOIL CLASS3 TIER 1 PCL (PPB)	MDL (PPB)	PQL (PPB)
Ethylene Dibromide (EDB) 1,2-Dibromoethane	106-93-4	Soil	21	0.01	0.03
1,2-Dibromo-3-Chloropropane (DCBP)	96-12-8	Soil	170	0.1	0.3
1,2,3-Trichloropropane	96-18-4	Soil	120	0.03	0.1

Note: The TRRP Tier 1 protective concentration levels (PCLs) are the default cleanup standards in the Texas Risk Reduction Program. The TCEQ has posted the updated TRRP Tier 1 Protective Concentration Levels (PCL) tables to the TRRP PCL [web page](#).






#4
 1,2-Dibromoethane
 Concen: 0.02 ug/L
 RT: 6.889 min Scan# 965
 Delta R.T. -0.017 min
 Lab File: 12051707.D
 Acq: 29 Apr 2012 2:52 am




Tgt Ion:107 Resp: 881

Ion	Ratio	Lower	Upper
107	100		
109	87.0	75.0	112.4








#10
 1,2-Dibromo-3-chloropr...
 Concen: 0.10 ug/L
 RT: 10.485 min Scan# 1688
 Delta R.T. 0.008 min
 Lab File: 12051721.D
 Acq: 29 Apr 2012 8:47 am




Tgt Ion: 75 Resp: 811

Ion	Ratio	Lower	Upper
75	100		
157	75.2	74.3	111.5
155	51.3	58.6	88.0#





#7
 1,2,3-Trichloropropane
 Concen: 0.04 ug/L
 RT: 8.462 min Scan# 1281
 Delta R.T. 0.002 min
 Lab File: 12051714.D
 Acq: 29 Apr 2012 5:49 am



Tgt Ion:110 Resp: 547

Ion	Ratio	Lower	Upper
110	100		
75	260.8	241.0	361.4
112	75.1	46.1	69.1#

